

Clinical Evaluation for

Embrace No Code Blood Glucose Monitoring System

Applicant: Apex Biotechnology Corp.

Test Date: 2007/8/28~2007/9/21

Clinical Site: SCHMIDT Group Practice Clinic
No. 29, Chung-shan Rd., Hsinchu, Taiwan, ROC
Tel: 03-5277333 Fax: 03-5277889

Investigator: Fu-Hsiang Chen Date: Sep-28-2007
MD, Department of
Gastroenterology and Internal Medicine

Co-investigator: May-Tze Lee Date: Sep-28-2007
MD, Department of
Endocrine and Metabolism

Co-investigator: Lih-Mei Chen Date: Sep-28-2007
MD, Department of
Neurology and Internal Medicine

1. Scope and Objective

The **primary objective** of this study is to verify the performance of testing on Embrace No Code system as compared to an accepted standard method of blood glucose analysis (the Yellow Springs Instrument [YSI]).

The assessment of each user's ability could understand and perform the operation of the systems while obtaining accurate results, and to assess the ease to be understanding of the User's manual, ease of use of the devices, and the acceptance degree of the voice feature.

2. Materials

- (1) Embrace No Code Blood Glucose Monitoring System:
 - Meters (6 sets): S/N TK0000000014 ~ 19
 - Test strips: 3 lots, lot No. FS002i, FS002J, FS002K, at least 500 pieces (10 vials) each lot
 - Embrace level 1 & 2 Control Solutions
 - User Manual
- (2) YSI 2300 Glucose Analyzer, reagents, and controls
- (3) Linearity control check solution: an NIST traceable NERL 1343 Standard Glucose Solution
- (4) Capi-ject capillary blood collection tubes
- (5) Micro-centrifuge for HCT detection
- (6) Lancing devices and Lancets
- (7) Case Report Forms, Data Logs and Questionnaires ([Appendix 1](#))
 - Form A: Informed Consent and Subject Information
 - Form B: YSI and Meters Quality Control Log
 - Form C: Test Data Log
 - Form D~F: Questionnaires: User Manual Questionnaire, Ease-of-Use Questionnaire and Voice Feature Questionnaire

3. Subjects

- (1) Total of 152 volunteers whose HCT is in the range of 30~55% participated in the study. The only exclusion criteria were bleeding diathesis or coagulation disease.
- (2) All volunteers filled the Subject Information and signed Consent before

testing.

- (3) The age distribution is 20 - 81 years old and the sex percentages are around 50% to 50%.
- (4) Eighty two (82) diabetics are included, twenty (20) are type I and sixty two (62) are type II. Forty five (45) diabetics are impaired visual, including the severe presbyopia, retinopathy, cataract or glaucoma. Except forty seven (47) diabetics, most participants are without the experience of performing self-monitoring blood glucose.
- (5) Education levels are junior high school to post graduate, and with at least 3 years English experience. The averaged English learning experience is 8.3 years. All the participants can read and speak English. They operated the Meter and performed blood glucose tests after reading the English User Manual.

4. Evaluation Site and Personnel

Investigation was conducted at SCHMIDT Group Practice Clinic, Hsinchu, Taiwan, directed by three clinicians. The project director is majored in Gastroenterology and Internal Medicine. Both co-investigators are majored in Endocrine and Metabolism and Neurology and Internal Medicine. The investigator responded to evaluate the qualified participants, answer to any questions regarding the study and control the proper participants numbers of different glucose levels.

Three trained technicians were also involved, two from ApexBio and one from SCHMIDT clinics to run YSI analyzer, meters tests and data collection. Healthcare Professional Tests were performed by the physicians or trained technicians.

5. Method

5.1 Device-Familiarization Period

The participants were familiar with the meter operation and sample adsorption after reading the Embrace No Code User's Guide. Only the English instructions and questionnaires were provided on study. The participants were confirmed be able to read and understand the English language. After the familiarization period, the participants performed the Control Solution (Level 1

and Level 2) tests to verify their operating techniques. Record the test readings of Control Solution and check with the Control Ranges labeled on the strip bottle.

5.2 Sample Collection & Meter testing

- (1) Participant washed hands with warm, soapy water or alcohol wipes. Hands are thoroughly rinsed and dried.
- (2) Participants performed finger stick to obtain Capi-ject collection for duplicate measurements with YSI reference analyzer. At least 250 μ L should be collected. Duplicate YSI measurement should take place within 10 minutes of collection.
- (3) With no assistance or guidance (other than from Participant's own reading of User's Manual), the Participant:

- Conducted self-test on Embrace No Code meter #1.

- (4) Healthcare Professional conducted:

- Embrace No Code meter test with Meter #1.
- Embrace No Code meter test with Meter #2.

Healthcare Professional recorded the time of both the tests of Participant and Professional and recorded observations, such as Participant errors or error messages requiring repeat self-testing.

- (5) Second finger stick performed by Professional to obtain Capi-ject collection for duplicate measurements with YSI reference analyzer and hematocrit determination. At least 250 μ L should be collected.

Healthcare professional made additional finger sticks as needed to obtain blood for YSI or hematocrit determination.

Healthcare professional interrogated meter memories and compared results from meter memory to results on Test Data Log. Study Coordinator indicated memory was checked on Test Data Log and recorded any discrepancies.

- (6) Alternate site testing (AST)

- Participant conducts self-test on palm.
- Professional conducts palm sampling.
- Participant conducts self-test on forearm.
- Professional conducts forearm sampling.

Participants performed alternate site testing started from palm or forearm

according to the advice of Study Coordinator. One half of Participants started AST testing with palm, ½ with forearm. One half of palm sampling was done on thenar, ½ on hypothenar. Professional tested same part of palm as Participant. Healthcare professional kept a running log to control the proper numbers of test sites started.

- (7) Third finger stick performed by Professional to obtain Capi-ject collection for duplicate measurements with YSI reference analyzer. At least 250 µL should be collected.
- (8) The Professional documented all blood glucose results on Test Data Log.
- (9) Professional collected additional one tube (EDTA anticoagulant) of finger stick capillary blood from some participants with low glucose level. These blood samples were incubated for about 3-4 hours allow glucose hydrolyzed to adequate low levels (< 50mg/dL), which were tested by professional only.

5.3 User Questionnaires

After the meters testing, participants refer to User Manual Sections that are referenced in the questionnaires while completing the following questionnaires.

- User Manual Questionnaire
- Ease-of-Use Questionnaire
- Voice Feature Questionnaire.

6. Measurement Procedures

6.1 Quality Control Check of YSI 2300 Analyzer

Daily quality control for the linearity checks of YSI 2300 analyzer was performed and recorded before testing and every two hours during testing period.

The NERL 1343 Standard Glucose Solutions were used for this study. Glucose levels of the NERL standards are 50 mg/dL, 100 mg/dL, 200 mg/dL, 400 mg/dL and 750 mg/dL.

Accuracy – mean results of duplicate tests should be within 5 mg/dL of the target glucose value at levels < 100 mg/dL or within 5% of the target glucose value at levels \geq 100 mg/dL.

6.2 Quality performance confirm for Embrace No Code meters

The Embrace No Code systems (meters and strips) were run the Control Solution tests to confirm their performance each day before testing.

Test result should be within the control ranges labeled on the strip bottle.

- 6.3 Finger stick capillary blood collection: after hand cleaning, each participant performed a finger stick by himself.
- 6.4 Record the testing time by healthcare professional.
- 6.5 The first capillary blood sample for YSI analyzer was collected and measured by healthcare professional. Record the sampling time.
- 6.6 Test on Embrace No Code meter following the User's Guide by the participant himself.
- 6.7 Test the same capillary sample on another 2 Embrace No Code meters by healthcare professional.
- 6.8 After the meters testing, perform a second fingerstick and collect the second fingerstick capillary blood sample for YSI analyzer. Record the sampling time. The time period for both YSI samples collected should not be longer than 10 minutes.
- 6.9 Collect a blood sample into a centrifuge with 12000 rpm for 5 minutes to measure the hematocrit (HCT).
- 6.10 The participant cleaned and dried through the palm and forearm.
- 6.11 According to the Healthcare Professional, Participant performed sticks on either palm or forearm in first and tested on meter, and then the other site.
- 6.12 Using the same stick, healthcare professional tested the AST sample on meter.
- 6.13 Healthcare Professional recorded the stick time and the measuring results of Participants and Professionals on the Test Data Logs.
- 6.14 After finishing the blood test, the participants were required to fill up designed Questionnaires to evaluate their understanding the User's Manual and the satisfactory of the system.

7. Statistical Analysis

Participant Information sheets are reviewed by Study Coordinators to determine

if the subject is taking a medication that may yield false results with the Embrace No Code system. Only results from Participants who have hematocrits between 30% and 55% will be used in determining the Embrace No Code system's performance. Data from pilot Participants will not be analyzed.

Steady state evaluation

Calculate the change percentage between the first and second fingerstick YSI results. Participants with changes larger than 10% are lack of steady state. Exclude non-steady state results for the System Accuracy Evaluation.

Calculate the change percentage between the second and third fingerstick YSI results to determine steady state for AST testing. Exclude any participant's outside 10% limits for steady state.

Test serial of palm and forearm and sampling site of palm

Count the number of subjects started from palm and count the number of puncture on Thenar.

7.1 Accuracy Evaluation

7.1.1 Correlation

Compare averaged duplicate YSI of first fingerstick vs. professional's both tests on meters for the System Accuracy Evaluation. The professional's meters value will be plotted on the vertical axis against the mean of the two YSI values on the horizontal axis. Using linear regression, the data of the individual Embrace No Code meter results versus YSI results will be analyzed. The slope, y-intercept and correlation coefficient will be noted on the scatter plot.

7.1.2 Bias analysis

(1) Calculate bias (mg/dL) if glucose level < 75 mg/dL or %bias if glucose level \geq 75 mg/dL between Embrace No Code meter's readings and YSI 2300.

(2) Use readings from YSI 2300 as X-axis and the calculated bias to be Y-axis. Plot the bias with +/- 20% as the control limits at glucose level higher than 75mg/dL and +/- 15mg/dL at glucose level lower than 75mg/dL.

(3) Count the numbers outside the acceptable control limits.

7.1.3 Error Grid Analysis

Count the numbers fall within Zone A~Zone E. 100% of the test data of Embrace No Code must be fall within zone A and B.

- (1) Zone A – glucose values deviating from the reference method by less than 20%
- (2) Zone B – deviation greater than 20%, leading to benign or no treatment error
- (3) Zone C – resulting in over-correcting acceptable blood glucose levels.
- (4) Zone D – dangerous failure to detection and treatment.
- (5) Zone E – erroneous treatment error.

7.2 User performance evaluation

7.2.1 Control Solution Test Evaluation

Control Solution test results from the users are recorded and compare with the Control Ranges on Strip labels to evaluate the user can get correct control solution results.

7.2.2 Blood Glucose Value Comparison

Blood glucose value obtained from fingerstick of untrained user is compared to results obtained from YSI Analyzer and fingerstick results obtained from professional. Calculate and compare the followings:

- (1) Compare participant's self-test vs. Professional's first meter test
- (2) Compare participant's self-test vs. averaged duplicate YSI of first fingerstick

7.3 For AST Evaluation

7.3.2 AST vs. Fingerstick Evaluation

Blood glucose values obtained form alternate sampling sites (palm and forearm) are compared to fingerstick results obtained from physician. Plot the results of forearm and palm on the vertical axis against the fingerstick results on the horizontal axis. Perform the linear regression analysis, note the slope, y-intercept and correlation coefficient on the scatter plots. Compare the followings:

- (1) Compare Professional's first meter test on fingerstick vs. Participant self-test on palm

- (2) Compare Professional's first meter test on fingerstick vs. Professional's test on palm
- (3) Compare Professional's first meter test on fingerstick vs. Participant self-test on forearm
- (4) Compare Professional's first meter test on fingerstick vs. Professional's test on forearm

7.4 User Questionnaires Evaluation

Data obtained from the three Questionnaires will be summarized and presented with the data analysis. The mean of the numerical responses for each statement and overall mean will be calculated and plotted as bar charts. Users' suggestions and comments will be summarized and discussed.

8. Results

8.1 Data Sheet

Total of 158 blood samples were tested, including 152 fresh capillary blood samples meet the hemmatocrit criteria of 30~55% and additional 6 glucose hydrolyzed incubated blood samples.

Change percentages of the 152 fresh samples are during -8% ~ 7% between the first and second fingerstick YSI results and -9% ~ 5% between the second and third YSI results. They meet in the steady state blood glucose criteria of within +/- 10%.

All the results from the qualified 158 samples are included in the meters accuracy evaluation by Healthcare Professional and only the 152 fresh blood samples are included in User and AST performance evaluations.

Three strip lots were used for testing of 158 samples randomly and almost equally during the study period. Forty nine (49) of blood samples were tested with strip lot FS002i, fifty sixty (56) with lot FS002J and fifty three (53) with lot FS002K.

For AST testing, half (76) of 152 participants started from palm, another half started from forearm. Half punctured palm on Thiner and another half on Hypothenar.

Re Appendix 2 for the Clinical Evaluation Worksheets

8.2 Accuracy Evaluation

8.2.1 Correlation between Embrace No Code against YSI

Re Exhibit 1 of Appendix 3 for the Correlation of Embrace No Code against YSI. The slope, y-intercept and correlation coefficient of linear regression are summarized on the table and noted on the scatter plot.

8.2.2 Bias Analysis

Re Exhibit 2 of Appendix 3 for the Bias Counts and Bias Plot of Embrace No Code against YSI. 100% of the 158 test samples fall within the acceptance criteria.

8.2.3 Clark Error Grid Analysis

Re Exhibit 3 of Appendix 3 for the Error Grid Analysis of Embrace No Code against YSI. 100% of the 158 test samples fall within Zone A, which is the acceptance area.

8.3 User Performance

8.3.1 Control Solution Test Evaluation

Re Exhibit 1 of Appendix 4 for the Control Solution Test evaluated by Lay Users. Total of 152 untrained lay users perform the correct Control Solution Tests. 100% of the test results of both the Level 1 and Level 2 Control Solution fall within the control ranges labeled on strip label. All the 152 users could perform the correct Control Solution tests after reading the User's Manual.

8.3.2 Blood Glucose Value Comparison

Re Exhibit 2 of Appendix 4 for the Comparison of participant's self-test vs. Professional's first meter test and Exhibit 3 of Appendix 4 for the Comparison of participant's self-test vs. averaged duplicate YSI of first fingerstick.

The slope, y-intercept and correlation coefficient of linear regression are summarized on the table and noted on the scatter plot.

8.4 For AST Evaluation

Re Exhibit 1 ~ 4 of Appendix 5, the followings are evaluated. The slope, y-intercept and correlation coefficient of linear regression are summarized on

the tables and noted on the scatter plots.

- Compare Professional's first meter test on fingerstick vs. Participant self-test on palm
- Compare Professional's first meter test on fingerstick vs. Professional's test on palm
- Compare Professional's first meter test on fingerstick vs. Participant self-test on forearm
- Compare Professional's first meter test on fingerstick vs. Professional's test on forearm

8.5 User Questionnaires Evaluation

8.5.1 Utility and Clarity of User's Guide

Re Exhibit 1 of Appendix 6 for the averaged rating score of the questionnaires for Utility and Clarity of User's Guide. Scores for all the 12 questions are averaged over 4. Most the users understand the indication on the User's Guide easily.

8.5.2 Ease of Use of Embrace No Code

Re Exhibit 2 of Appendix 6 for the averaged rating score for Ease of Use of Embrace No Code. Scores for all the 12 questions are averaged of 4.3 ~ 4.7. Design and performance of Embrace No Code are acceptable by most the users. The lowest score items are the setting the meter and perform the AST sampling.

8.5.2 Acceptance Degree of Voice Feature

Re Exhibit 3 of Appendix 6 for the averaged rating score for Acceptance Degree of Voice Feature. Scores for all the 5 questions are all over 4.5. Users with impaired visual agree that the accessory voice function is user friendly, to assistant for testing. Users with normal sight like the voice function because of the smart and innovation design.

9. Conclusion

According to the clinical study, as compared to the reference method YSI 2300 analyzer, the GlycoSure Voice show good correlation, 100% of the 158

test samples fall within the acceptance criteria of bias analysis, and 100% fall within the Zone A of Clark Error Grid analysis.

For the AST tests, both the trained professional and untrained user can get the accurate results, correlated well to YSI. According to the clear indication on the User's Guide, users can perform the correct sampling and testing to generate the accurate results.

After reading the User's Manual, even the untrained lay users can operate the Embrace No Code meter correctly to get the accurate blood glucose value, correlated to trained healthcare professional.

Most users understand the indication on User's Guide and can operate the meter without assistants. The most satisfied feature of the new system is the Voice function, because it helps to remind for the test results.

10. REFERENCE

- 10.1 FDA, ICH Harmonized Tripartite Guideline, Guideline for Good Clinical Practice, 1 May 1996.
- 10.2 ISO 15197, In vitro diagnostic test system – Requirements for blood-glucose monitoring systems for self-testing in managing diabetes mellitus. First edition, May 2003.
- 10.3 AdvaMed Draft Guidance Document – Review Criteria for Assessment of Portable Blood Glucose Monitoring In Vitro Diabetic Devices.
- 10.4 Draft Guidance for Industry and FDA Staff - Total Product Life Cycle for Portable Invasive Blood Glucose Monitoring Systems -- APPENDIX 1. ALTERNATIVE SAMPLE SITE TESTING (AST).
- 10.5 Clarke WL, Cox D, Gonder-Frederick LA, Carter W, Pohl SL: Evaluating clinical accuracy of systems for self-monitoring of blood glucose. Diabetes Care 10:622–628, 1987.

Appendix 2

Clinical Evaluation Worksheets

Meter: Embrace No Code
 Lot no: FS002i, FS002J, FS002K

Test Date:8/28/~9/21/2007
 Temp.oC: 23.1 ~25.5
 RH%: 45 ~ 65 %

Test No.	HCT	Fingertip-YSI Test						Change (%)		Fingertip			Palm		Forearm	
		YSI-1st Finger Stick		YSI-2nd Finger Stick		YSI-3rd Finger Stick		1st vs. 2nd	2nd vs. 1st	User	Professional		User	Profess ional	User	Profes sional
		Test 1	Test 2	Test 1	Test 2	Test 1	Test 2			Meter 1	Meter 1	Meter 2				
1	41	333	334	333	333	333	334	0%	0%	340	344	347	349	349	351	347
2	43	214	214	214	211	213	214	-1%	0%	222	223	210	218	216	194	188
3	49	234	233	234	233	234	234	0%	0%	215	226	226	207	203	188	219
4	37	457	457	454	456	457	459	0%	1%	424	422	443	472	447	423	461
5	45	48.8	46	44.7	44.1	46	45	-6%	2%	65	62	55	62	62	58	60
6	45	389	386	385	384	386	386	-1%	0%	367	372	382	372	382	358	370
7	45	272	270	271	268	270	271	-1%	0%	243	251	262	248	261	281	274
8	41	194	196	197	195	195	197	1%	0%	182	187	198	209	226	178	201
9	40	290	288	291	288	289	290	0%	0%	292	263	305	324	310	270	274
10	48	222	219	225	223	223	222	2%	-1%	208	198	201	186	221	198	229
11	42	366	364	363	362	364	364	-1%	0%	389	378	367	364	386	344	325
12	43	208	206	210	210	209	208	1%	-1%	194	199	210	199	228	200	218
13	43	234	233	241	241	238	237	3%	-1%	219	211	213	206	208	222	245
14	43	319	317	331	332	326	324	4%	-2%	326	343	318	334	373	294	353
15	36	362	363	364	361	362	364	0%	0%	370	352	405	337	413	362	418

Test No.	HCT	Fingertip-YSI Test						Change (%)		Fingertip			Palm		Forearm	
		YSI-1st Finger Stick		YSI-2nd Finger Stick		YSI-3rd Finger Stick		1st vs. 2nd	2nd vs. 1st	User	Professional		User	Profess ional	User	Profes sional
		Test 1	Test 2	Test 1	Test 2	Test 1	Test 2			Meter 1	Meter 1	Meter 2				
16	48	417	417	426	424	420	421	2%	-1%	399	408	407	389	429	383	424
17	43	215	215	215	214	215	215	0%	0%	210	211	229	203	254	224	220
18	45	245	245	243	242	244	244	-1%	1%	237	238	222	245	259	239	248
19	45	270	268	270	268	269	269	0%	0%	249	262	261	272	253	258	260
20	38	356	354	358	354	355	356	0%	0%	368	373	377	348	360	362	348
21	47	413	416	413	414	414	415	0%	0%	403	392	420	382	409	390	400
22	36	405	403	406	407	406	405	1%	0%	429	411	393	414	428	420	432
23	43	253	252	256	253	253	254	1%	0%	242	228	254	268	259	251	234
24	43	51.2	49.3	52.9	51.6	51.4	51.1	4%	-2%	56	50	57	54	50	60	62
25	43	451	449	453	452	451	451	1%	0%	456	475	457	424	450	410	464
26	38	386	385	387	387	387	386	0%	0%	358	407	393	370	384	379	376
27	49	245	244	244	243	244	244	0%	0%	240	228	228	213	240	226	242
28	35	478	476	479	479	478	478	0%	0%	491	499	504	473	487	465	468
29	38	196	198	197	195	196	198	-1%	1%	201	194	211	184	192	189	188
30	38	254	253	261	257	256	257	2%	-1%	257	277	270	232	248	248	252
31	37	320	318	319	317	319	319	0%	0%	308	325	331	330	310	339	323
32	44	253	251	254	252	253	253	0%	0%	249	241	252	269	269	226	228
33	44	196	193	195	192	194	194	-1%	0%	192	204	189	195	208	203	223

Test No.	HCT	Fingertip-YSI Test						Change (%)		Fingertip			Palm		Forearm	
		YSI-1st Finger Stick		YSI-2nd Finger Stick		YSI-3rd Finger Stick		1st vs. 2nd	2nd vs. 1st	User	Professional		User	Professional	User	Professional
		Test 1	Test 2	Test 1	Test 2	Test 1	Test 2			Meter 1	Meter 1	Meter 2				
34	52	315	316	317	314	315	317	0%	0%	292	289	299	283	296	268	312
35	42	352	353	354	353	353	354	0%	0%	373	343	371	364	352	335	341
36	40	377	379	378	378	378	379	0%	0%	376	364	370	376	389	388	385
37	41	231	228	235	234	233	232	2%	-1%	233	227	238	244	230	221	241
38	42	368	367	375	375	371	371	2%	-1%	361	378	389	374	383	387	376
39	41	274	273	274	275	274	274	0%	0%	261	271	269	251	278	246	267
40	39	174	173	176	176	175	174	1%	-1%	187	171	174	166	168	169	175
41	36	470	472	472	471	471	472	0%	0%	509	503	498	480	451	505	457
42	43	182	181	181	182	182	181	0%	0%	170	175	175	178	170	178	176
43	37	277	276	275	276	276	276	0%	0%	292	275	267	276	276	248	269
44	44	191	194	194	194	192	194	1%	-1%	176	192	187	178	189	179	179
45	41	170	168	165	163	167	167	-3%	2%	170	160	156	149	166	161	159
46	43	163	162	163	162	163	163	0%	0%	163	149	162	154	153	140	146
47	46	183	182	184	183	183	183	1%	0%	166	176	171	170	168	169	175
48	49	177	176	181	179	178	179	2%	-1%	171	173	159	163	193	166	160
49	43	139	138	138	136	138	138	-1%	1%	134	139	148	128	143	123	154
50	39	123	122	123	122	123	123	0%	0%	135	119	126	131	124	144	123
51	43	262	261	263	259	261	262	0%	0%	256	251	250	242	263	246	259

Test No.	HCT	Fingertip-YSI Test						Change (%)		Fingertip			Palm		Forearm	
		YSI-1st Finger Stick		YSI-2nd Finger Stick		YSI-3rd Finger Stick		1st vs. 2nd	2nd vs. 1st	User	Professional		User	Professional	User	Professional
		Test 1	Test 2	Test 1	Test 2	Test 1	Test 2			Meter 1	Meter 1	Meter 2				
52	42	275	273	278	273	274	276	1%	0%	279	259	280	247	254	256	290
53	42	159	158	157	156	158	158	-1%	1%	161	139	152	166	151	155	165
54	39	181	181	184	183	182	183	1%	-1%	171	169	167	190	181	186	187
55	43	301	302	299	299	301	300	-1%	1%	293	291	290	291	300	303	303
56	39	178	179	178	177	178	178	-1%	0%	171	175	179	182	166	188	186
57	38	164	165	168	164	165	166	1%	0%	176	179	168	165	180	164	167
58	37	74.7	75.1	78.4	76.3	75.7	76.6	3%	-2%	85	78	85	84	87	88	80
59	48	188	189	180	179	184	184	-5%	3%	197	202	205	207	199	201	199
60	42	79.5	79.9	80.7	79.6	79.8	80.1	1%	0%	81	85	75	70	75	76	77
61	42	91.9	92.4	93.8	91.6	92	92.9	1%	0%	83	92	90	86	79	89	81
62	48	186	187	188	184	185	187	0%	0%	172	169	181	186	176	161	174
63	41	99.6	101	105	104	103	102	4%	-2%	85	87	90	90	90	85	95
64	43	240	241	240	241	241	240	0%	0%	248	267	270	238	262	219	256
65	50	177	178	181	180	191	179	2%	2%	183	179	185	175	180	184	174
66	42	190	191	191	191	191	191	0%	0%	193	215	205	195	195	182	181
67	46	81	81.5	82	80.2	80.9	81.5	0%	0%	95	94	92	84	81	91	83
68	48	74.1	74.6	75.3	73	73.8	74.7	0%	0%	67	80	81	69	81	62	84
69	46	314	315	315	313	314	315	0%	0%	327	313	302	309	314	287	309

Test No.	HCT	Fingertip-YSI Test						Change (%)		Fingertip			Palm		Forearm	
		YSI-1st Finger Stick		YSI-2nd Finger Stick		YSI-3rd Finger Stick		1st vs. 2nd	2nd vs. 1st	User	Professional		User	Professional	User	Professional
		Test 1	Test 2	Test 1	Test 2	Test 1	Test 2			Meter 1	Meter 1	Meter 2				
70	46	73.7	76.6	81.2	78.3	77.5	77.5	6%	-3%	88	82	86	78	88	85	83
71	52	139	139	132	133	136	136	-5%	3%	126	124	124	119	119	115	123
72	42	139	141	133	136	138	137	-4%	2%	119	125	120	141	128	131	124
73	45	113	114	112	111	113	113	-2%	1%	107	106	110	114	109	113	95
74	46	86.9	87.7	86.3	86.5	87.1	86.6	-1%	1%	90	92	102	104	98	92	96
75	45	106	106	106	106	106	106	0%	0%	97	93	89	94	95	92	93
76	42	88.2	88.3	87.7	87.6	87.9	88	-1%	0%	88	82	88	93	90	76	86
77	45	196	196	194	194	195	195	-1%	1%	198	191	191	204	196	178	183
78	38	82.4	83.3	82.5	82.3	82.8	82.5	-1%	0%	86	82	77	89	84	82	77
79	42	82.1	80.8	79.4	78.9	79.9	80.8	-3%	2%	82	70	81	85	72	76	76
80	41	174	174	186	187	171	170	7%	-9%	181	178	163	171	177	185	181
81	48	92.9	91.2	88.8	87.3	89.3	90.9	-4%	2%	96	88	82	85	89	79	97
82	44	136	134	131	129	123	123	-4%	-5%	124	138	131	135	129	130	127
83	47	102	101	96.4	97.3	94.2	94.2	-5%	-3%	98	97	85	85	86	90	91
84	42	92.7	91	88.8	88	89.5	90.8	-4%	2%	89	87	78	87	77	75	82
85	43	90.9	87.5	89.8	87.7	87.6	90.3	-1%	0%	81	82	93	80	76	87	78
86	41	288	287	284	285	286	286	-1%	1%	300	294	287	293	291	292	296
87	49	87	85.9	83.4	84.4	80.2	80.2	-3%	-4%	88	81	77	72	86	70	83

Test No.	HCT	Fingertip-YSI Test						Change (%)		Fingertip			Palm		Forearm	
		YSI-1st Finger Stick		YSI-2nd Finger Stick		YSI-3rd Finger Stick		1st vs. 2nd	2nd vs. 1st	User	Professional		User	Professional	User	Professional
		Test 1	Test 2	Test 1	Test 2	Test 1	Test 2			Meter 1	Meter 1	Meter 2				
88	41	144	143	143	142	143	143	-1%	0%	134	128	133	150	144	142	136
89	44	164	164	162	162	163	163	-1%	1%	159	156	141	141	169	137	161
90	53	181	180	185	185	183	183	2%	-1%	162	172	175	152	177	168	167
91	44	73.9	73.7	76	76.1	79.9	80	3%	5%	68	66	76	75	68	70	65
92	47	155	154	154	152	154	154	-1%	1%	168	160	160	169	147	159	159
93	45	138	136	139	137	138	138	1%	0%	129	152	137	137	136	131	139
94	51	142	142	139	139	141	141	-2%	1%	142	137	151	117	140	124	136
95	44	183	185	184	183	183	185	0%	0%	184	172	177	169	176	178	189
96	40	178	178	175	176	177	177	-1%	1%	175	166	161	166	167	173	159
97	44	82.4	81.6	84.4	84.3	83	83.4	3%	-1%	85	82	75	82	84	77	91
98	42	177	176	179	179	178	178	1%	-1%	168	169	163	171	179	177	173
99	47	178	175	179	177	178	177	1%	0%	181	181	165	178	174	169	189
100	50	161	160	162	162	162	161	1%	0%	162	158	160	158	164	147	170
101	45	78.8	78.7	72.9	74.6	76.7	75.9	-6%	3%	71	72	69	68	80	65	79
102	42	79.3	78.4	80.7	80.9	79.7	80	2%	-1%	76	77	86	77	76	76	81
103	47	385	383	377	376	380	380	-2%	1%	371	372	377	380	375	366	370
104	48	75.3	74.2	77.4	77.2	75.7	76.3	3%	-2%	82	75	84	76	79	64	81
105	45	189	189	194	194	191	192	3%	-1%	170	170	169	179	166	166	158

Test No.	HCT	Fingertip-YSI Test						Change (%)		Fingertip			Palm		Forearm	
		YSI-1st Finger Stick		YSI-2nd Finger Stick		YSI-3rd Finger Stick		1st vs. 2nd	2nd vs. 1st	User	Professional		User	Profess ional	User	Profes sional
		Test 1	Test 2	Test 1	Test 2	Test 1	Test 2			Meter 1	Meter 1	Meter 2				
106	45	74.2	73.5	73.1	72.3	72.9	73.7	-2%	1%	79	78	83	73	88	68	82
107	43	452	450	445	445	447	448	-1%	1%	474	431	426	460	483	443	481
108	46	261	260	266	265	263	263	2%	-1%	262	257	260	256	281	238	276
109	45	137	136	140	140	139	138	3%	-1%	127	153	135	133	139	116	141
110	44	148	147	145	144	146	146	-2%	1%	147	159	141	156	147	137	142
111	48	187	185	185	184	186	185	-1%	1%	193	172	181	184	185	190	188
112	41	195	194	193	192	193	194	-1%	1%	194	192	198	185	200	190	185
113	41	137	135	138	136	137	137	1%	0%	128	125	126	146	136	133	137
114	47	148	147	148	148	148	148	0%	0%	127	149	151	131	138	149	142
115	47	151	152	152	151	151	151	0%	0%	161	158	145	140	150	136	146
116	43	76.7	76	75.9	76.1	76.1	76.3	0%	0%	82	79	84	70	74	62	75
117	45	77.6	77.2	76.5	75.7	76.5	77.1	-2%	1%	82	86	76	87	88	76	81
118	47	127	126	129	127	127	128	1%	0%	117	141	121	114	121	138	139
119	47	84	83.6	78.5	77.9	80.8	81.3	-7%	4%	80	78	82	74	90	83	83
120	46	79.7	78.6	73.1	72	75.3	76.4	-8%	5%	81	83	80	83	83	76	90
121	43	170	169	166	166	167	168	-2%	1%	161	170	179	180	174	175	176
122	46	75.4	75.1	77.6	77.8	76.5	76.5	3%	-2%	75	76	82	77	77	79	78
123	42	132	129	131	129	129	131	0%	0%	130	145	136	140	130	127	130

Test No.	HCT	Fingertip-YSI Test						Change (%)		Fingertip			Palm		Forearm	
		YSI-1st Finger Stick		YSI-2nd Finger Stick		YSI-3rd Finger Stick		1st vs. 2nd	2nd vs. 1st	User	Professional		User	Professional	User	Professional
		Test 1	Test 2	Test 1	Test 2	Test 1	Test 2			Meter 1	Meter 1	Meter 2				
124	44	74	72.9	77.5	76.1	79.5	80.8	5%	4%	81	82	80	84	78	82	72
125	43	101	99	98.5	97.7	98.4	100	-2%	1%	110	114	111	107	106	117	95
126	43	77.8	77.1	79.2	78.2	77.7	78.5	2%	-1%	89	85	89	87	90	90	88
127	48	87.3	87.5	82.6	82.9	85.2	85	-5%	3%	79	85	81	78	92	79	77
128	42	94	93.9	94.7	93.6	93.8	94.5	0%	0%	102	101	105	100	88	98	91
129	46	92.7	93.8	94.4	94.8	94.3	93.6	1%	-1%	92	96	91	86	88	85	83
130	42	83.7	83.6	85.5	84.2	83.9	84.6	1%	-1%	92	87	84	95	95	85	92
131	40	79.6	80.6	77.4	79.2	79.9	78.5	-2%	1%	78	77	80	70	72	73	73
132	44	90.6	91.4	90.2	90.7	91.1	90.4	-1%	0%	76	77	76	78	81	78	79
133	42	83.9	83.5	86.6	88	85.8	85.3	4%	-2%	97	90	90	82	86	87	78
134	40	76.7	78.7	78	78.3	78.5	77.4	1%	0%	81	85	91	91	84	93	91
135	41	65.6	66.4	66.2	66.3	66.4	65.9	0%	0%	74	73	74	79	74	76	71
136	44	80.7	81.7	80.8	80.7	81.2	80.8	-1%	0%	68	75	77	80	89	76	77
137	44	201	203	199	197	199	200	-2%	1%	196	197	198	201	194	190	197
138	44	91.4	92.2	90.5	91	91.6	91	-1%	1%	93	100	103	97	85	96	92
139	41	116	118	117	117	118	117	0%	0%	99	107	98	107	108	102	105
140	48	84.5	86.1	82.7	82.3	84.2	83.6	-3%	2%	85	84	89	74	80	71	77
141	49	74.9	74.8	77.4	75.9	75.4	76.2	2%	-1%	82	79	79	85	82	80	74

Test No.	HCT	Fingertip-YSI Test						Change (%)		Fingertip			Palm		Forearm	
		YSI-1st Finger Stick		YSI-2nd Finger Stick		YSI-3rd Finger Stick		1st vs. 2nd	2nd vs. 1st	User	Professional		User	Professional	User	Professional
		Test 1	Test 2	Test 1	Test 2	Test 1	Test 2			Meter 1	Meter 1	Meter 2				
142	41	78.1	77.8	80	78.3	78.1	79.6	2%	0%	82	83	84	88	79	80	76
143	47	78.9	78.3	77.7	77.6	78	78.3	-1%	1%	75	80	80	66	84	78	76
144	43	81.2	81.2	77.7	79.2	80.2	79.5	-3%	2%	92	95	82	89	83	91	88
145	45	81.1	83.4	83.3	83.2	83.3	82.2	1%	-1%	79	78	79	81	81	76	75
146	41	90.9	91.2	93.6	92	91.6	92.3	2%	-1%	96	103	101	95	96	100	105
147	46	74.1	74.2	72.3	71.8	73	73.2	-3%	1%	80	80	86	87	86	78	77
148	45	85.2	83.4	83.4	82	82.7	84.3	-2%	1%	71	81	89	78	83	80	77
149	41	251	252	255	253	252	253	1%	-1%	248	255	251	222	252	224	244
150	44	79.2	77.1	79.7	78.2	72.7	74.5	1%	-7%	86	83	80	80	80	78	85
151	42	90.2	90.8	91.4	90.5	90.7	90.8	0%	0%	79	81	85	78	79	82	78
152	41	80.3	81	80.9	80.4	80.7	80.6	0%	0%	83	86	93	71	79	90	78
153*		44.2	43.6								52	54				
154*		39.6	41.2								48	45				
155*		47.5	47.6								53	57				
156*		41.3	40.5								50	53				
157*		47.6	48								52	59				
158*		34.3	35.8								46	49				

*: Glucose hydrolyzed samples for blood glucose level of <50mg/dL.

Exhibit 1 of Appendix 3

Correlation of Embrace No Code vs. YSI

Accuracy - Correlation

Linear Regression Data:

Embrace No Code Meter vs. YSI 2300 PLUS STAT analyzer			
<i>n</i>	Slope	Intercept	r value
158	0.996	0.644 mg/dL	0.995

Scatter Plot:




Exhibit 2 of Appendix 3

Bias Analysis of
Embrace No Code vs. YSI

Accuracy – Bias Analysis

Bias Counts:

Bias	Embrace No Code	
	No.	Ratio (%)
Glucose level ≥ 75 mg/dL		
< 5%	87	55%
5-10%	40	25%
10-15%	16	10%
15-20%	1	0.6%
Sum of ≥ 75 mg/dL	144	91%
Glucose level < 75 mg/dL		
< 15mg/dL	14	9%
> 15mg/dL	0	0%
Sum of < 75 mg/dL	14	9%
Total	158	100%

Acceptance criteria:

For glucose level ≥ 75 mg/dL: +/- 20% of YSI

For glucose level < 75 mg/dL: +/- 15mmg/dL

Bias Plot:




Exhibit 3 of Appendix 3

Clark Error Grid Analysis of
Embrace No Code vs. YSI

Accuracy - Clark Error Grid Plot




Exhibit 1 of Appendix 4

Control Solution Test Results

Evaluation of Lay User

Evaluation of Control Solution Test Results of Lay Users

Control Solution Test Result of Users:

	Level 1	Level 2
Readings of tests	62~88 mg/dL	138~189mg/dL
# of within range	152	152
Percentage	100%	100%

Participant No.: 152

Lot # of test strip: FS002E

Control Range - Level 1: 59~89mg/dL

- Level 2: 138~206mg/dL

Lot # of control solution:

- Level 1: TC960707M

- Level 2: TC960701H

Exhibit 2 of Appendix 4

Comparison of Participant's
Self-test vs. Professional's Test

Comparison of Participant's Self-test vs. Professional's Test - Fingertip

Linear Regression Data:

User vs. Professional of Fingertip			
<i>n</i>	Slope	Intercept	r value
152	1.002	-0.445 mg/dL	0.994

Scatter Plot of User vs. Professional:




Exhibit 3 of Appendix 4

Comparison of User's Self-test vs. YSI

Comparison of Participant's Self-test vs. YSI - Fingertip

Linear Regression Data:

User's Fingertip vs. YSI Analyzer			
<i>n</i>	Slope	Intercept	r value
152	1.000	-1.381 mg/dL	0.994

Scatter Plot:




Exhibit 1 of Appendix 5

Comparison of Professional's
Fingerstick vs. Participant's Palm

Comparison of Participant's Self-Palm Test vs. Professional's Fingertip Test

Linear Regression Data:

User's Palm vs. Professional's Fingertip			
<i>n</i>	Slope	Intercept	r value
152	0.981	1.530 mg/dL	0.990

Scatter Plot:




Exhibit 2 of Appendix 5

Comparison of Professional's
Fingerstick vs. Professional's Palm

Comparison of Palm Test vs. Fingertip Test of Professional

Linear Regression Data:

Palm vs. Fingertip of Professional			
<i>n</i>	Slope	Intercept	r value
152	1.015	-0.623 mg/dL	0.993

Scatter Plot:




Exhibit 3 of Appendix 5

Comparison of Professional's
Fingerstick vs. Participant's Forearm

Comparison of Participants' Forearm Test vs. Fingertip Test of Professional

Linear Regression Data:

User's Forearm vs. Professional's Fingertip			
<i>n</i>	Slope	Intercept	r value
152	0.968	0.668 mg/dL	0.991

Scatter Plot:




Exhibit 4 of Appendix 5

Comparison of Professional's
Fingerstick vs. Professional's Forearm

Comparison of Palm Test vs. Fingertip Test of Professional

Linear Regression Data:

Forearm vs. Fingertip of Professional			
<i>n</i>	Slope	Intercept	r value
152	1.002	-0.334 mg/dL	0.992

Scatter Plot:




Exhibit 1 of Appendix 6

Utility and Clarity of User's Manual

Questionnaire for Utility and Clarity of User's Guide




Exhibit 2 of Appendix 6

Ease of Use of Embrace No Code

Questionnaire for Ease of Use of Embrace No Code




Exhibit 3 of Appendix 6

Voice Feature of Embrace No Code

Questionnaire for Acceptance Degree of Voice Feature

